



Technical specifications

Intended use	Conservative and Restorative Dentistry					
Indications for use	Minimal excavation and macro-abrasion					
Description	Diamond-coated, left-angled half-spherical tip to access the lesion without damaging the adjacent tooth					
Tip type	Diamond-coated					
Allergenic factor	Tip made of surgical stainless steel as per ASTM F899 - 12b					
	Surgical stainless steel contains: <ul style="list-style-type: none">• nickel• chromium					
	Diamond polishing crimping binder alloy contains: <ul style="list-style-type: none">• nickel					
Commercial reference	EXL : F02044					
Medical class	Class of medical device: IIa according to 93/42/EEC directive					
Adjustments	Newtron range	P5 Range	Suprasson P5 Booster	Prophy Max - P-Max - Implant Center 2 - Piezotome 2		Irrigation
	12	12	9	2		

Service life

As it is not possible to establish a maximum number of uses (that may be determined by many parameters such as duration of use, hardness of tissue, the force applied, wear), we recommend that medical devices being routinely used are replaced at least once a year. Replace the medical device if the tip does not vibrate at the expected frequency, the treatment is not progressing as normal and is taking longer or is at a standstill.


Associated documentation

Document title	References
Consulting electronic user instructions	J00007
Cleaning, disinfection and sterilisation instructions for tips	J02001
Excavus Kit clinical tips	J08141
Ultrasonic generator power settings table	J58000
General instructions relating to the complete range of standard ultrasonic dental tips	J02101


Sterilisation



Sterilise prior to each use




The tips must be cleaned, disinfected and sterilised prior to each use. Prior to each use, please refer to the cleaning, disinfection and sterilisation instructions for tips J02001.



Installation



Disassembly



Manufacturer identification



SATELEC
A Company of ACTEON Group
17, avenue Gustave Eiffel
ZI du Phare
33700 MERIGNAC
France
Tel. +33 (0) 556.34.06.07
Fax. +33 (0) 556.34.92.92
E.mail : satelec@acteongroup.com
www.acteongroup.com



Date of first CE marking

1998



The practitioner has to continually monitor the clinical site to immediately identify any heating risk.

Tip | EXL | J00581 | (98) | V3 | 05/2021 | NB26EN040C

SATELEC | A Company of ACTEON Group
17 av. Gustave Eiffel | ZI du Phare
33700 MERIGNAC | FRANCE
Tel. +33 (0) 556 34 06 07 | Fax. +33 (0) 556 34 92 92
satelec@acteongroup.com | www.acteongroup.com

Download



Instructions For Use

